

QIAGEN Supplementary Protocol

QuantiTect® Probe RT-PCR Kit research protocol for S-OIV

This protocol is for use in swine-origin influenza A (H1N1) virus (S-OIV) research applications using sequences available from the World Health Organization (WHO) (www.who.int/csr/resources/publications/swineflu/realtimeptcr/en/). Real-time one-step RT-PCR is performed on the Applied Biosystems® 7500 or other real-time PCR cycler using the QuantiTect Probe RT-PCR Kit in combination with degenerate primers and probe.

IMPORTANT: Please consult the “Safety Information” and “Important Notes” sections in the *QuantiTect Probe RT-PCR Handbook* before beginning this procedure. For safety information on the additional chemicals mentioned in this protocol, please consult the appropriate material safety data sheets (MSDSs), available from the product supplier.

The QuantiTect Probe RT-PCR Kit is intended for research use. No claim or representation is intended to provide information for the diagnosis, prevention, or treatment of a disease.

Equipment and reagents

When working with chemicals, always wear a suitable lab coat, disposable gloves, and protective goggles. For more information, consult the appropriate material safety data sheets (MSDSs), available from the product supplier.

- QuantiTect Probe RT-PCR Kit (cat. no. 204443 or 204445)
- Primers and probes: These should be purchased from an established oligonucleotide manufacturer. Primers should be of standard quality, and probes should be HPLC purified. Lyophilized primers and probes should be dissolved in TE buffer to provide a stock solution of 100 μ M; concentration should be checked by spectrophotometry (for details, visit www.qiagen.com/resources/info/Guidelines_RTPCR/Assay_Probe.aspx). Primer and probe stock solutions should be stored in aliquots at -20°C . Probe stock solutions should be protected from exposure to light.
- Nuclease-free (RNase/DNase free) consumables: Special care should be taken to avoid nuclease contamination of all reagents and consumables used to set up PCR for sensitive detection of viral nucleic acids.
- Cooling device or ice
- Applied Biosystems 7500 Real-Time PCR System or other real-time PCR cycler
- PCR tubes or plates (use thin-walled PCR tubes or plates recommended by the manufacturer of the cycler)



- Optional: Trizma® base and EDTA for preparing TE buffer for storing primers and probes (see www.qiagen.com/resources/info/Guidelines_RTPCR/Assay_Probe.aspx). Use RNase/DNase-free water and plastic consumables to prepare TE buffer.

Important points before starting

- Always start with the cycling conditions specified in this protocol.
- After reverse transcription, the PCR step of the RT-PCR must start with an **initial incubation step of 15 min at 95°C** to activate HotStarTaq® DNA Polymerase.
- Set up all reactions on ice to avoid premature cDNA synthesis.
- Always readjust the threshold value for analysis of every run.
- 2x QuantiTect Probe RT-PCR Master Mix contains dUTP, which allows the use of a uracil-N-glycosylase (UNG) pretreatment of the reaction. Only **heat-labile** UNG should be used.
- When using the **Applied Biosystems 7500**, it is necessary to adjust the preset threshold value to a lower value. Use a value of 0.01 as a starting point.

Procedure

1. **Thaw 2x QuantiTect Probe RT-PCR Master Mix (if stored at –20°C), template RNA, primer and probe solutions, and RNase-free water. Mix the individual solutions, and place them on ice. QuantiTect RT Mix should be taken from –20°C immediately before use, always kept on ice, and returned to storage at –20°C immediately after use.**
2. **Prepare a reaction mix according to Table 1.**

Keep samples on ice while preparing the reaction mix.

Note: We strongly recommend starting with an initial Mg^{2+} concentration of 4 mM as provided by 2x QuantiTect Probe RT-PCR Master Mix. For a few targets, reactions may be improved by using Mg^{2+} concentrations of up to 6 mM.

Table 1. Reaction setup

Component	Volume per 25 μ l reaction	Final concentration
2x QuantiTect Probe RT-PCR Master Mix*	12.5 μ l	1x
Forward primer	Variable	0.8 μ M
Reverse primer	Variable	0.8 μ M
Probe	Variable	0.2 μ M [†]
QuantiTect RT Mix	0.25 μ l	–
Template RNA (added at step 4)	5 μ l	–
RNase-free water	Variable	–
Optional: Uracil-N-glycosylase, heat-labile	Variable	2 units/reaction
Total volume per reaction	25 μl	–

* Provides a final concentration of 4 mM MgCl₂.

[†] A final probe concentration of 0.2 μ M gives satisfactory results when using the InfA, SW InfA, and RNase P primer–probe sets. **For the SW H1 primer–probe set, use a final probe concentration of 0.4 μ M.**

3. Mix the reaction mix thoroughly, and dispense appropriate volumes into PCR tubes or plates.

Keep the PCR tubes or plates on ice.

4. Add template RNA (5 μ l) to the individual PCR tubes or wells containing the reaction mix.

5. Program your real-time cycler according to the program outlined in Table 2.

For optional UNG treatment, leave the samples for at least 5 min on ice.

6. Place the PCR tubes or plates in the real-time cycler, and start the cycling program.

Table 2. Cycling conditions

Step	Time	Temperature	Additional comments
Reverse transcription	30 min	50°C	Temperatures up to 55°C can be used to eliminate secondary structure in the template RNA.
PCR initial activation step	15 min	95°C	HotStarTaq DNA Polymerase is activated by this heating step.
2-step cycling:			
Denaturation	15 s	95°C	
Combined annealing/extension	60 s	52°C	Perform fluorescence data collection
Number of cycles	45		

QIAGEN handbooks can be requested from QIAGEN Technical Service or your local QIAGEN distributor. Selected handbooks can be downloaded from www.qiagen.com/literature.

Material safety data sheets (MSDS) for any QIAGEN product can be downloaded from www.qiagen.com/Support/MSDS.aspx.

Trademarks: QIAGEN®, HotStarTaq®, QuantiTect® (QIAGEN Group); Applied Biosystems® (Applied Biosystems Corporation or its subsidiaries); Trizma® (Sigma-Aldrich Co.). Registered names, trademarks, etc. used in this document, even when not specifically marked as such, are not to be considered unprotected by law.

NOTICE TO PURCHASER: LIMITED LICENSE

A license to perform the 5' nuclease process for research requires the use of a Licensed 5' Nuclease Kit (containing Licensed Probe), or the combination of an Authorized 5' Nuclease Core Kit plus Licensed Probe, or license rights that may be purchased from Applied Biosystems. This product (QuantiTect Probe RT-PCR Kit) is an Authorized 5' Nuclease Core Kit without Licensed Probe. Its purchase price includes a limited, non-transferable immunity from suit under U.S. Patents Nos. 5,210,015, 5,487,972, 5,476,774, and 5,219,727, and corresponding patent claims outside the United States, owned by Roche Molecular Systems, Inc. or F. Hoffmann-La Roche Ltd (Roche), for using only this amount of the product in the practice of the 5' nuclease process solely for the purchaser's own internal research when used in conjunction with Licensed Probe. This product is also an Authorized 5' Nuclease Core Kit for use with service sublicenses available from Applied Biosystems. This product conveys no rights under U.S. Patents Nos. 5,804,375, 6,214,979, 5,538,848, 5,723,591, 5,876,930, 6,030,787, or 6,258,569, or corresponding patents outside the United States, expressly, by implication or by estoppel. No right under any other patent claims (such as apparatus or system claims in U.S. Patent No. 6,814,934) and no right to perform commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is hereby granted expressly, by implication or by estoppel. This product is for research use only. Diagnostic uses require a separate license from Roche. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

NOTICE TO PURCHASER: LIMITED LICENSE

The purchase price of this product (QuantiTect Probe RT-PCR Kit) includes a limited, non-transferable license under U.S. Patents Nos. 5,407,800, 5,322,770, 5,310,652 and corresponding patent claims outside the United States, owned by Roche Molecular Systems, Inc. or F. Hoffmann-La Roche Ltd (Roche), to use only this amount of product solely for the purchaser's own internal research. No right under any other patent claims (such as apparatus or system claims) and no right to use this product for any other purpose or for commercial services of any kind, including without limitation reporting the results of purchaser's activities for a fee or other commercial consideration, is hereby granted expressly, by implication or by estoppel. This product is for research use only. Diagnostic uses require a separate license from Roche. Further information on purchasing licenses may be obtained by contacting the Director of Licensing, Applied Biosystems, 850 Lincoln Centre Drive, Foster City, California 94404, USA.

NOTICE TO PURCHASER: LIMITED LICENSE

This product (QuantiTect Probe RT-PCR Kit) or its use is covered by at least one claim of U.S. Pat. Nos. 5,035,996; 5,945,313; 6,287,823; or 6,518,026, owned by Invitrogen Corporation. The purchase of this product conveys to the buyer the non-transferable right to use the purchased amount of the product and components of the product in research conducted by the buyer (whether the buyer is an academic or for-profit entity). The buyer cannot sell or otherwise transfer (a) this product, (b) its components, or (c) materials made by the employment of this product or its components to a third party or otherwise use this product or its components or materials made by the employment of this product or its components for Commercial Purposes. Commercial Purposes means any activity for which a party receives or is due to receive consideration and may include, but is not limited to: (1) use of the product or its components in manufacturing; (2) use of the product or its components to provide a service, information, or data; (3) use of the product or its components for therapeutic, diagnostic or prophylactic purposes; or (4) resale of the product or its components, whether or not such product or its components are resold for use in research. The buyer cannot use this product or its components or materials made using this product or its components for therapeutic, diagnostic or prophylactic purposes. Further information on purchasing licenses under the above patents may be obtained by contacting the Licensing Department, Invitrogen Corporation, 1600 Faraday Avenue, Carlsbad, CA 92008. Email: outlicensing@invitrogen.com.

PCR111 Aug-09 © 2009 QIAGEN, all rights reserved.

www.qiagen.com

Australia = 1-800-243-800

Austria = 0800/281010

Belgium = 0800-79612

Canada = 800-572-9613

China = 021 3865 3865

Denmark = 80-885945

Finland = 0800-914416

France = 01-60-920-930

Germany = 02103-29-12000

Hong Kong = 800 933 965

Ireland = 1800 555 049

Italy = 800 787980

Japan = 03-5547-0811

Korea (South) = 1544 7145

Luxembourg = 8002 2076

The Netherlands = 0800 0229592

Norway = 800-18859

Singapore = 65-67775366

Spain = 91-630-7050

Sweden = 020-790282

Switzerland = 055-254-22-11

UK = 01293-422-911

USA = 800-426-8157

